

CERTIFICATE

It is hereby confirmed that the

Bench Fume Cupboard 1500 x 900 x 2400

of the manufacturer

ENLAB INDUSTRIAL LABORATORY SYSTEMS and EQUIPMENTS

was tested according to

EN 14175 part 3 and 6

at the total volume flow rate of **1150** m³/h for constant volume flow rate and
200 - 1150 m³/h with Schneider VAV system FC 500
with the nominal face velocity of **0.5** m/s

The details and data of the complete test procedure
are content of the type test report

FCS/EN/252/2017

which is given to the manufacturer on

3rd February 2017

Tintschl BioEnergie und Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen
Germany

Controlled by:

.....
Dipl.-Ing. Bernd Schubert

Tested by:

.....
Dipl.-Ing. Zheni Uzunova
Dipl.-Chem. Sulay Tröger



Erlangen, 02 / 02 / 2017